


A16_03_5K

Company:
 User:
 Result file: C:\CamsizerXT\CAMDAT\A16_03_5K\A16_03_5K049.rdf
 Task file: C:\CamsizerXT\CAMSYS\A16_03_5K.afg
 Time: 7/31/2015, 11:58, duration 41 min 16 s at 0.5 % covered area, image rate 1:1, with X-Fall, gap width = 7.0 mm

Particle model: xc_min
 No. of particles: CCD-B = 97813473, CCD-Z = 1315998
 Velocity adaption: no
 Fitting: no

Material:

ASTM(+)	No(-)	retained [%]	passing [%]	SPHT3	Symm3	b/l3	PDN
> #16		0.0	100.0				0
#20	#16	0.0	100.0	0.884	0.899	0.722	1
#30	#20	0.0	100.0				0
#40	#30	0.0	100.0	0.773	0.857	0.671	14
#50	#40	0.0	100.0	0.776	0.827	0.725	55
#60	#50	0.4	100.0	0.890	0.884	0.790	2447
#70	#60	4.2	99.6	0.918	0.902	0.782	48130
#140	#70	64.0	95.4	0.917	0.909	0.760	3173762
PAN	#140	31.4	31.4	0.897	0.920	0.737	258570466



Mean value SPHT3 = 0.904
 Mean value Symm3 = 0.911
 Mean value b/l3 = 0.739
 Mean value RDNS_C3 = 0.667
 Mean value SPHT_K3 = 0.678

User