


A19\_SPHERE\_AND\_ROUND

Company:  
 User:  
 Result file: C:\CamsizerXT\CAMDATA\A19\_SPHERE\_AND\_ROUND\A19\_SPHERE\_AND\_ROUND250.rdf  
 Task file: C:\CamsizerXT\CAMSYS\A19\_SPHERE\_AND\_ROUND.afg  
 Time: 7/13/2015, 15:23, duration 10 min 57 s at 0.5 % covered area, image rate 1:1, with X-Fall, gap width = 7.0 mm

Particle model: xc\_min  
 No. of particles: CCD-B = 15545529, CCD-Z = 184017  
 Velocity adaption: no  
 Fitting: no

Material:

ASTM(+)	No(-)	retained [%]	passing [%]	SPHT3	Symm3	b/l3	PDN
> #16		0.0	100.0				0
#20	#16	0.0	100.0				0
#30	#20	0.0	100.0				0
#40	#30	0.0	100.0	0.330	0.744	0.681	1
#50	#40	0.0	100.0	0.594	0.720	0.702	75
#60	#50	0.8	100.0	0.827	0.844	0.757	2655
#70	#60	8.5	99.2	0.888	0.883	0.764	45194
#140	#70	83.9	90.7	0.902	0.902	0.748	1696710
PAN	#140	6.8	6.8	0.878	0.918	0.691	32019522



Mean value SPHT3 = 0.893  
 Mean value Symm3 = 0.899  
 Mean value b/l3 = 0.732  
 Mean value RDNS\_C3 = 0.602  
 Mean value SPHT\_K3 = 0.670

User