


A7\_04\_SPHERE\_ROUND

Company:	
User:	
Result file:	C:\CamsizerXT\CAMDAT\A7_04_SPHERE_ROUND\A7_04_SPHERE_ROUND144.rdf
Task file:	C:\CamsizerXT\CAMSYS\A7_04_SPHERE_ROUND.afg
Time:	6/16/2015 , 9:48 , duration 6 min 14 s at 0.5 % covered area, image rate 1:1 , with X-Fall, gap width = 7.0 mm
Particle model:	xc_min
No. of particles:	CCD-B = 5256721 , CCD-Z = 58853
Velocity adaption:	no
Fitting:	no
Material:	

ASTM(+)	No(-)	retained [%]	passing [%]	SPHT3	Symm3	b/l3	PDN
> #16		0.1	100.0	0.542	0.755	0.514	1
#20	#16	0.0	99.9	0.487	0.826	0.393	2
#30	#20	0.1	99.9	0.760	0.893	0.832	1
#40	#30	0.0	99.8	0.726	0.844	0.626	7
#50	#40	0.0	99.8	0.576	0.713	0.691	45
#60	#50	0.3	99.8	0.788	0.817	0.748	949
#70	#60	3.4	99.5	0.888	0.879	0.772	16321
#140	#70	90.5	96.1	0.907	0.906	0.745	1551421
PAN	#140	5.6	5.6	0.862	0.918	0.655	11949493



Mean value SPHT3 = 0.897
Mean value Symm3 = 0.903
Mean value b/l3 = 0.727
Mean value RDNS_C3 = 0.598
Mean value SPHT_K3 = 0.655

User