


A71_SPHERE_ROUND

Company:	
User:	
Result file:	C:\CamsizerXT\CAMDAT\A71_SPHERE_ROUND\A71_SPHERE_ROUND139.rdf
Task file:	C:\CamsizerXT\CAMSYS\A7-1_SPHERE_ROUND.afg
Time:	6/15/2015 , 15:36 , duration 5 min 23 s at 0.5 % covered area, image rate 1:1 , with X-Fall, gap width = 7.0 mm
Particle model:	xc_min
No. of particles:	CCD-B = 4753288 , CCD-Z = 118134
Velocity adaption:	no
Fitting:	no
Material:	

ASTM(+)	No(-)	retained [%]	passing [%]	SPHT3	Symm3	b/l3	PDN
> #16		0.0	100.0				0
#20	#16	0.0	100.0				0
#30	#20	0.0	100.0				0
#40	#30	0.0	100.0	0.978	0.958	0.802	1
#50	#40	0.1	100.0	0.671	0.730	0.716	242
#60	#50	2.2	99.9	0.877	0.871	0.767	9660
#70	#60	17.7	97.7	0.916	0.900	0.767	116412
#140	#70	78.9	80.0	0.911	0.907	0.732	1340115
PAN	#140	1.1	1.1	0.849	0.921	0.630	24071137



Mean value SPHT3 = 0.904
Mean value Symm3 = 0.903
Mean value b/l3 = 0.722
Mean value RDNS_C3 = 0.633
Mean value SPHT_K3 = 0.641

User