


B50_SPHER_ROUND

Company:	
User:	
Result file:	C:\CamsizerXT\CAMDAT\B50_SPHER_ROUND\B50_SPHER_ROUND073.rdf
Task file:	C:\CamsizerXT\CAMSYS\B50_SPHER_ROUND.afg
Time:	9/25/2015 , 13:47 , duration 8 min 8 s at 0.5 % covered area, image rate 1:1 , with X-Fall, gap width = 7.0 mm
Particle model:	xc_min
No. of particles:	CCD-B = 1738561 , CCD-Z = 46048
Velocity adaption:	no
Fitting:	no
Material:	

ASTM(+)	No(-)	retained [%]	passing [%]	SPHT3	Symm3	b/l3	PDN
> #16		0.0	100.0				0
#20	#16	0.0	100.0				0
#30	#20	0.0	100.0				0
#40	#30	0.0	100.0				0
#50	#40	0.1	100.0	0.687	0.753	0.729	208
#60	#50	2.3	99.9	0.871	0.868	0.766	9527
#70	#60	16.3	97.6	0.903	0.891	0.760	103136
#140	#70	80.6	81.3	0.899	0.902	0.719	1229564
PAN	#140	0.7	0.7	0.830	0.915	0.594	5833568



Mean value SPHT3 = 0.891
Mean value Symm3 = 0.898
Mean value b/l3 = 0.710
Mean value RDNS_C3 = 0.575
Mean value SPHT_K3 = 0.614

User