


B56_01_PRECRUSH

Company:
 User:
 Result file: C:\CamsizerXT\CAMDAT\B56_01_PRECRUSH\B56_01_PRECRUSH077.rdf
 Task file: C:\CamsizerXT\CAMSYS\B56_01_PRECRUSH.afg
 Time: 9/17/2015 , 15:26 , duration 14 min 53 s at 0.5 % covered area, image rate 1:1 , with X-Fall, gap width = 7.0 mm

Particle model: xc_min
 No. of particles: CCD-B = 9833144 , CCD-Z = 205121
 Velocity adaption: no
 Fitting: no

Material:

ASTM(+)	No(-)	retained [%]	passing [%]	SPHT3	Symm3	b/l3	PDN
> #16		0.0	100.0				0
#20	#16	0.0	100.0	0.833	0.916	0.767	2
#30	#20	0.3	100.0	0.831	0.896	0.783	46
#40	#30	0.6	99.7	0.830	0.883	0.765	365
#50	#40	2.6	99.1	0.861	0.880	0.769	4287
#60	#50	5.1	96.5	0.890	0.887	0.766	16378
#70	#60	9.4	91.4	0.904	0.893	0.760	48079
#140	#70	70.4	82.0	0.909	0.907	0.740	1320087
PAN	#140	11.6	11.6	0.888	0.928	0.712	49361054



Mean value SPHT3 = 0.898
 Mean value Symm3 = 0.904
 Mean value b/l3 = 0.727
 Mean value RDNS_C3 = 0.648
 Mean value SPHT_K3 = 0.654

User