


B59_1_SNR

Company:	
User:	
Result file:	C:\CamsizerXT\CAMDAT\B59_1_SNR\B59_1_SNR129.rdf
Task file:	C:\CamsizerXT\CAMSYS\B59_1_SNR.afg
Time:	10/20/2015 , 13:27 , duration 8 min 7 s at 0.5 % covered area, image rate 1:1 , with X-Fall, gap width = 7.0 mm
Particle model:	xc_min
No. of particles:	CCD-B = 1580415 , CCD-Z = 19201
Velocity adaption:	no
Fitting:	no
Material:	

ASTM(+)	No(-)	retained [%]	passing [%]	SPHT3	Symm3	b/l3	PDN
> #16		0.0	100.0				0
#20	#16	0.0	100.0				0
#30	#20	0.0	100.0				0
#40	#30	0.0	100.0				0
#50	#40	0.0	100.0	0.652	0.741	0.731	51
#60	#50	1.0	100.0	0.851	0.858	0.767	2154
#70	#60	10.0	99.0	0.903	0.893	0.770	36314
#140	#70	79.1	89.0	0.902	0.903	0.739	1000754
PAN	#140	9.9	9.9	0.886	0.922	0.695	2325135



Mean value SPHT3 = 0.894
Mean value Symm3 = 0.902
Mean value b/l3 = 0.724
Mean value RDNS_C3 = 0.660
Mean value SPHT_K3 = 0.647

User