


DILL\_01\_SV\_14\_CRUSH6K

Company:  
 User:  
 Result file: C:\CamsizerXT\CAMDAT\DILL\_01\_SV\_14\_CRUSH6K\DILL\_01\_SV\_14\_CRUSH6K027.rdf  
 Task file: C:\CamsizerXT\CAMSYS\DILL\_01\_SV\_14\_CRUSH6K.afg  
 Time: 2/5/2015 , 12:51 , duration 25 min 58 s at 0.5 % covered area, image rate 1:1 , with X-Fall, gap width = 7.0 mm

Particle model: xc\_min  
 No. of particles: CCD-B = 139340706 , CCD-Z = 1108248  
 Velocity adaption: no  
 Fitting: no

Material:

ASTM(+)	No(-)	retained [%]	passing [%]	SPHT3	Symm3	b/l3	PDN
> #16		0.0	100.0				0
#20	#16	0.0	100.0				0
#30	#20	0.0	100.0				0
#40	#30	0.0	100.0				0
#50	#40	0.0	100.0	0.703	0.772	0.735	202
#60	#50	0.9	100.0	0.890	0.880	0.785	14425
#70	#60	8.8	99.1	0.922	0.902	0.780	214615
#140	#70	76.4	90.3	0.914	0.905	0.746	6290527
PAN	#140	13.9	13.9	0.866	0.904	0.694	252389914



Mean value SPHT3 = 0.902  
 Mean value Symm3 = 0.903  
 Mean value b/l3 = 0.727  
 Mean value RDNS\_C3 = 0.581  
 Mean value SPHT\_K3 = 0.654

User