


DILL\_01\_SV\_14\_SandR\_140MESH

Company:  
 User:  
 Result file: C:\...izerXT\CAMDAT\DILL\_01\_SV\_14\_SandR\_140MESH\DILL\_01\_SV\_14\_SandR\_140MESH023.rdf  
 Task file: C:\CamsizerXT\CAMSYS\DILL\_01\_SV\_14\_SandR\_140MESH.afg  
 Time: 2/3/2015 , 14:21 , duration 10 min 9 s at 0.5 % covered area, image rate 1:1 , with X-Fall, gap width = 7.0 mm

Particle model: xc\_min  
 No. of particles: CCD-B = 2366570 , CCD-Z = 24366  
 Velocity adaption: no  
 Fitting: no

Material:

ASTM(+)	No(-)	retained [%]	passing [%]	SPHT3	Symm3	b/l3	PDN
> #16		0.0	100.0				0
#20	#16	0.0	100.0				0
#30	#20	0.0	100.0				0
#40	#30	0.0	100.0				0
#50	#40	0.1	100.0	0.609	0.711	0.705	111
#60	#50	1.0	99.9	0.846	0.850	0.761	3685
#70	#60	9.7	98.9	0.905	0.890	0.768	53298
#140	#70	85.4	89.2	0.901	0.901	0.735	1552397
PAN	#140	3.8	3.8	0.846	0.913	0.633	4913468



Mean value SPHT3 = 0.892  
 Mean value Symm3 = 0.898  
 Mean value b/l3 = 0.721  
 Mean value RDNS\_C3 = 0.590  
 Mean value SPHT\_K3 = 0.640

User