


DILL_02_SV_14_SPH_ROUN

Company:	
User:	
Result file:	C:\CamsizerXT\CAMDAT\DILL_02_SV_14_SPH_ROUN\DILL_02_SV_14_SPH_ROUN026.rdf
Task file:	C:\CamsizerXT\CAMSYS\DILL_02_SV_14_SPH_ROUN.afg
Time:	2/6/2015 , 9:35 , duration 5 min 5 s at 0.5 % covered area, image rate 1:1 , with X-Fall, gap width = 7.0 mm
Particle model:	xc_min
No. of particles:	CCD-B = 2770502 , CCD-Z = 34576
Velocity adaption:	no
Fitting:	no
Material:	

ASTM(+)	No(-)	retained [%]	passing [%]	SPHT3	Symm3	b/l3	PDN
> #16		0.0	100.0				0
#20	#16	0.0	100.0				0
#30	#20	0.0	100.0				0
#40	#30	0.0	100.0	0.492	0.593	0.706	33
#50	#40	2.7	100.0	0.857	0.855	0.756	7917
#60	#50	39.4	97.3	0.922	0.906	0.773	206275
#70	#60	45.8	57.9	0.922	0.912	0.737	326091
#140	#70	12.0	12.1	0.887	0.906	0.641	119747
PAN	#140	0.1	0.1	0.841	0.893	0.684	8867582



Mean value SPHT3 = 0.911
Mean value Symm3 = 0.905
Mean value b/l3 = 0.728
Mean value RDNS_C3 = 0.627
Mean value SPHT_K3 = 0.654

User