


DILL_04_SV_14_SANDR_70

Company:
 User:
 Result file: C:\CamsizerXT\CAMDAT\DILL_04_SV_14_SANDR_70\DILL_04_SV_14_SANDR_70023.rdf
 Task file: C:\CamsizerXT\CAMSYS\DILL_04_SV_14_SANDR_70.afg
 Time: 2/9/2015 , 15:56 , duration 5 min 48 s at 0.5 % covered area, image rate 1:1 , with X-Fall, gap width = 7.0 mm

Particle model: xc_min
 No. of particles: CCD-B = 2346613 , CCD-Z = 64315
 Velocity adaption: no
 Fitting: no

Material:

ASTM(+)	No(-)	retained [%]	passing [%]	SPHT3	Symm3	b/l3	PDN
> #16		0.0	100.0				0
#20	#16	0.0	100.0				0
#30	#20	0.0	100.0				0
#40	#30	0.0	100.0	0.589	0.681	0.709	25
#50	#40	2.6	100.0	0.838	0.843	0.750	7329
#60	#50	37.6	97.4	0.913	0.900	0.771	192375
#70	#60	46.8	59.8	0.919	0.909	0.742	328458
#140	#70	13.0	13.0	0.890	0.906	0.650	128962
PAN	#140	0.0	0.0	0.870	0.889	0.717	5643351



Mean value SPHT3 = 0.904
 Mean value Symm3 = 0.900
 Mean value b/l3 = 0.727
 Mean value RDNS_C3 = 0.605
 Mean value SPHT_K3 = 0.657

User