


DILL_05_SV_14_SPHERE_ROUND

Company:	
User:	
Result file:	C:\...zerXT\CAMDAT\DILL_05_SV_14_SPHERE_ROUND\DILL_05_SV_14_SPHERE_ROUND023.rdf
Task file:	C:\CamsizerXT\CAMSYS\DILL_05_SV_14_SPHERE_ROUND.afg
Time:	8/10/2015 , 10:56 , duration 6 min 31 s at 0.5 % covered area, image rate 1:1 , with X-Fall, gap width = 7.0 mm
Particle model:	xc_min
No. of particles:	CCD-B = 4947262 , CCD-Z = 166752
Velocity adaption:	no
Fitting:	no
Material:	

ASTM(+)	No(-)	retained [%]	passing [%]	SPHT3	Symm3	b/l3	PDN
> #16		0.0	100.0				0
#20	#16	0.0	100.0				0
#30	#20	0.0	100.0				0
#40	#30	0.0	100.0				0
#50	#40	0.0	100.0	0.536	0.683	0.683	17
#60	#50	0.5	100.0	0.838	0.850	0.769	1058
#70	#60	5.3	99.5	0.890	0.884	0.769	16939
#140	#70	84.1	94.2	0.904	0.902	0.748	1109658
PAN	#140	10.1	10.1	0.899	0.928	0.706	25046730



Mean value SPHT3 = 0.896
Mean value Symm3 = 0.902
Mean value b/l3 = 0.731
Mean value RDNS_C3 = 0.651
Mean value SPHT_K3 = 0.663

User