


ENN_04_SV_14_SNR

Company:	
User:	
Result file:	C:\CamsizerXT\CAMDAT\ENN_04_SV_14_SNR\ENN_04_SV_14_SNR079.rdf
Task file:	C:\CamsizerXT\CAMSYS\ENN_04_SV_14_SNR.afg
Time:	6/1/2015 , 14:19 , duration 9 min 11 s at 0.5 % covered area, image rate 1:1 , with X-Fall, gap width = 7.0 mm
Particle model:	xc_min
No. of particles:	CCD-B = 1231892 , CCD-Z = 30447
Velocity adaption:	no
Fitting:	no
Material:	

ASTM(+)	No(-)	retained [%]	passing [%]	SPHT3	Symm3	b/l3	PDN
> #16		0.0	100.0				0
#20	#16	0.0	100.0				0
#30	#20	0.0	100.0				0
#40	#30	0.0	100.0				0
#50	#40	0.1	100.0	0.643	0.726	0.695	100
#60	#50	1.6	99.9	0.860	0.861	0.764	4046
#70	#60	16.2	98.3	0.910	0.897	0.767	61264
#140	#70	80.3	82.1	0.904	0.905	0.737	848395
PAN	#140	1.8	1.8	0.866	0.927	0.662	4200903



Mean value SPHT3 = 0.897
Mean value Symm3 = 0.902
Mean value b/l3 = 0.726
Mean value RDNS_C3 = 0.627
Mean value SPHT_K3 = 0.653

User