


FB\_02\_SV\_14\_SPHER\_ROUND\_70

Company:	
User:	
Result file:	C:\...zerXT\CAMDAT\FB_02_SV_14_SPHER_ROUND_70\FB_02_SV_14_SPHER_ROUND_70083.rdf
Task file:	C:\CamsizerXT\CAMSYS\FB_02_SV_14_SPHER_ROUND_70.afg
Time:	6/4/2015 , 13:22 , duration 10 min 54 s at 0.5 % covered area, image rate 1:1 , with X-Fall, gap width = 7.0 mm
Particle model:	xc_min
No. of particles:	CCD-B = 1239166 , CCD-Z = 19101
Velocity adaption:	no
Fitting:	no
Material:	

ASTM(+)	No(-)	retained [%]	passing [%]	SPHT3	Symm3	b/l3	PDN
> #16		0.0	100.0				0
#20	#16	0.0	100.0				0
#30	#20	0.0	100.0				0
#40	#30	0.0	100.0	0.551	0.639	0.768	9
#50	#40	3.2	100.0	0.860	0.861	0.760	8467
#60	#50	35.2	96.8	0.911	0.899	0.769	163676
#70	#60	45.0	61.6	0.912	0.905	0.734	289904
#140	#70	16.6	16.6	0.870	0.899	0.629	153989
PAN	#140	0.0	0.0	0.857	0.892	0.690	5985430



Mean value SPHT3 = 0.897
Mean value Symm3 = 0.899
Mean value b/l3 = 0.717
Mean value RDNS_C3 = 0.497
Mean value SPHT_K3 = 0.633

User