


## GFN-01\_7K

Company:	
User:	
Result file:	C:\CamsizerXT\CAMDAT\GFN-01_7K\GFN-01_7K034.rdf
Task file:	C:\CamsizerXT\CAMSYS\GFN-01_7K.afg
Time:	7/29/2015 , 9:59 , duration 31 min 46 s at 0.5 % covered area, image rate 1:1 , with X-Fall, gap width = 7.0 mm
Particle model:	xc_min
No. of particles:	CCD-B = 101759864 , CCD-Z = 755758
Velocity adaption:	no
Fitting:	no
Material:	

ASTM(+)	No(-)	retained [%]	passing [%]	SPHT3	Symm3	b/l3	PDN
> #16		0.0	100.0				0
#20	#16	0.0	100.0	0.908	0.929	0.818	2
#30	#20	0.0	100.0	0.808	0.889	0.678	7
#40	#30	0.0	100.0	0.846	0.890	0.715	18
#50	#40	0.1	100.0	0.779	0.819	0.724	295
#60	#50	1.4	99.9	0.897	0.888	0.783	15453
#70	#60	13.9	98.5	0.928	0.910	0.780	227869
#140	#70	71.7	84.6	0.918	0.911	0.753	3652070
PAN	#140	12.9	12.9	0.847	0.895	0.698	190917508



Mean value SPHT3 = 0.905
Mean value Symm3 = 0.908
Mean value b/l3 = 0.737
Mean value RDNS_C3 = 0.537
Mean value SPHT_K3 = 0.678

User