


GFN_01_6K

Company:	
User:	
Result file:	C:\CamsizerXT\CAMDAT\GFN_01_6K\GFN_01_6K031.rdf
Task file:	C:\CamsizerXT\CAMSYS\GFN_01_6K.afg
Time:	7/28/2015 , 12:18 , duration 30 min 34 s at 0.5 % covered area, image rate 1:1 , with X-Fall, gap width = 7.0 mm
Particle model:	xc_min
No. of particles:	CCD-B = 27172993 , CCD-Z = 587907
Velocity adaption:	no
Fitting:	no
Material:	

ASTM(+)	No(-)	retained [%]	passing [%]	SPHT3	Symm3	b/l3	PDN
> #16		0.0	100.0				0
#20	#16	0.0	100.0				0
#30	#20	0.0	100.0	0.886	0.893	0.699	1
#40	#30	0.0	100.0	0.854	0.877	0.746	2
#50	#40	0.0	100.0	0.707	0.764	0.730	208
#60	#50	1.3	100.0	0.892	0.884	0.780	12333
#70	#60	12.6	98.7	0.927	0.908	0.780	187272
#140	#70	76.8	86.1	0.919	0.913	0.753	3734371
PAN	#140	9.3	9.3	0.883	0.922	0.717	98666388



Mean value SPHT3 = 0.911
Mean value Symm3 = 0.911
Mean value b/l3 = 0.739
Mean value RDNS_C3 = 0.653
Mean value SPHT_K3 = 0.679

User