


GFN\_02\_SV\_14\_SNR

Company:	
User:	
Result file:	C:\CamsizerXT\CAMDAT\GFN_02_SV_14_SNR\GFN_02_SV_14_SNR078.rdf
Task file:	C:\CamsizerXT\CAMSYS\GFN_02_SV_14_SNR.afg
Time:	6/1/2015 , 13:50 , duration 9 min 39 s at 0.5 % covered area, image rate 1:1 , with X-Fall, gap width = 7.0 mm
Particle model:	xc_min
No. of particles:	CCD-B = 1109137 , CCD-Z = 26790
Velocity adaption:	no
Fitting:	no
Material:	

ASTM(+)	No(-)	retained [%]	passing [%]	SPHT3	Symm3	b/l3	PDN
> #16		0.0	100.0				0
#20	#16	0.0	100.0				0
#30	#20	0.0	100.0				0
#40	#30	0.0	100.0				0
#50	#40	0.0	100.0	0.587	0.715	0.716	23
#60	#50	0.5	100.0	0.774	0.825	0.746	624
#70	#60	6.6	99.5	0.857	0.872	0.756	13792
#140	#70	83.6	92.9	0.871	0.887	0.736	704410
PAN	#140	9.3	9.3	0.875	0.920	0.689	3938058



Mean value SPHT3 = 0.864
Mean value Symm3 = 0.887
Mean value b/l3 = 0.720
Mean value RDNS_C3 = 0.592
Mean value SPHT_K3 = 0.650

User