


GFS_06_SV_14_SPHERE_ROUND

Company:
 User:
 Result file: C:\...zerXT\CAMDAT\GFS_06_SV_14_SPHERE_ROUND\GFS_06_SV_14_SPHERE_ROUND176.rdf
 Task file: C:\CamsizerXT\CAMSYS\GFS_06_SV_14_SPHERE_ROUND.afg
 Time: 6/24/2015, 16:25, duration 5 min 30 s at 0.5 % covered area, image rate 1:1, with X-Fall, gap width = 7.0 mm

Particle model: xc_min
 No. of particles: CCD-B = 4163810, CCD-Z = 50685
 Velocity adaption: no
 Fitting: no

Material:

ASTM(+)	No(-)	retained [%]	passing [%]	SPHT3	Symm3	b/l3	PDN
> #16		0.0	100.0				0
#20	#16	0.0	100.0				0
#30	#20	0.0	100.0				0
#40	#30	0.0	100.0	0.570	0.663	0.711	14
#50	#40	3.2	100.0	0.856	0.859	0.759	10606
#60	#50	37.8	96.8	0.911	0.900	0.769	219508
#70	#60	45.3	59.0	0.915	0.908	0.737	364092
#140	#70	13.6	13.7	0.878	0.904	0.636	154323
PAN	#140	0.1	0.1	0.846	0.895	0.695	12871320



Mean value SPHT3 = 0.900
 Mean value Symm3 = 0.901
 Mean value b/l3 = 0.723
 Mean value RDNS_C3 = 0.533
 Mean value SPHT_K3 = 0.647

User