


GFS_14_SPHERE_ROUND

Company:
 User:
 Result file: C:\CamsizerXT\CAMDAT\GFS_14_SPHERE_ROUND\GFS_14_SPHERE_ROUND226.rdf
 Task file: C:\CamsizerXT\CAMSYS\GFS_14_SPHERE_ROUND.afg
 Time: 7/9/2015 , 10:23 , duration 5 min 52 s at 0.5 % covered area, image rate 1:1 , with X-Fall, gap width = 7.0 mm

Particle model: xc_min
 No. of particles: CCD-B = 3610362 , CCD-Z = 125153
 Velocity adaption: no
 Fitting: no

Material:

ASTM(+)	No(-)	retained [%]	passing [%]	SPHT3	Symm3	b/l3	PDN
> #16		0.0	100.0				0
#20	#16	0.0	100.0				0
#30	#20	0.0	100.0				0
#40	#30	0.0	100.0				0
#50	#40	0.1	100.0	0.661	0.723	0.717	193
#60	#50	2.0	99.9	0.879	0.872	0.769	9496
#70	#60	16.5	97.9	0.916	0.900	0.766	116041
#140	#70	80.0	81.4	0.917	0.912	0.738	1508584
PAN	#140	1.4	1.4	0.896	0.931	0.690	18972693



Mean value SPHT3 = 0.909
 Mean value Symm3 = 0.907
 Mean value b/l3 = 0.727
 Mean value RDNS_C3 = 0.683
 Mean value SPHT_K3 = 0.652

User