


GFS_18_SV_14_SPHERE_ROUND

Company:
 User:
 Result file: C:\...zerXT\CAMDAT\GFS_18_SV_14_SPHERE_ROUND\GFS_18_SV_14_SPHERE_ROUND186.rdf
 Task file: C:\CamsizerXT\CAMSYS\GFS_18_SV_14_SPHERE_ROUND.afg
 Time: 6/25/2015, 15:11, duration 5 min 12 s at 0.5 % covered area, image rate 1:1, with X-Fall, gap width = 7.0 mm

Particle model: xc_min
 No. of particles: CCD-B = 761740, CCD-Z = 19443
 Velocity adaption: no
 Fitting: no

Material:

ASTM(+)	No(-)	retained [%]	passing [%]	SPHT3	Symm3	b/l3	PDN
> #16		0.0	100.0				0
#20	#16	0.0	100.0				0
#30	#20	0.0	100.0	0.430	0.652	0.753	1
#40	#30	0.2	100.0	0.658	0.713	0.709	106
#50	#40	42.3	99.8	0.914	0.907	0.767	77082
#60	#50	49.1	57.5	0.914	0.913	0.729	129555
#70	#60	7.6	8.4	0.886	0.911	0.643	26090
#140	#70	0.7	0.8	0.827	0.901	0.542	3780
PAN	#140	0.1	0.1	0.867	0.881	0.702	10482182



Mean value SPHT3 = 0.905
 Mean value Symm3 = 0.908
 Mean value b/l3 = 0.725
 Mean value RDNS_C3 = 0.638
 Mean value SPHT_K3 = 0.650

User