


GFS\_20\_SPHERE\_ROUND

Company:	
User:	
Result file:	C:\CamsizerXT\CAMDAT\GFS_20_SPHERE_ROUND\GFS_20_SPHERE_ROUND238.rdf
Task file:	C:\CamsizerXT\CAMSYS\GFS_20_SPHERE_ROUND.afg
Time:	7/9/2015 , 14:23 , duration 6 min 10 s at 0.5 % covered area, image rate 1:1 , with X-Fall, gap width = 7.0 mm
Particle model:	xc_min
No. of particles:	CCD-B = 3280343 , CCD-Z = 103681
Velocity adaption:	no
Fitting:	no
Material:	

ASTM(+)	No(-)	retained [%]	passing [%]	SPHT3	Symm3	b/l3	PDN
> #16		0.0	100.0				0
#20	#16	0.0	100.0				0
#30	#20	0.0	100.0				0
#40	#30	0.0	100.0				0
#50	#40	0.1	100.0	0.682	0.733	0.716	327
#60	#50	2.8	99.9	0.883	0.876	0.769	13451
#70	#60	23.8	97.1	0.923	0.906	0.770	175715
#140	#70	72.4	73.3	0.918	0.911	0.737	1277501
PAN	#140	0.9	0.9	0.892	0.929	0.676	16982883



Mean value SPHT3 = 0.911
Mean value Symm3 = 0.907
Mean value b/l3 = 0.731
Mean value RDNS_C3 = 0.651
Mean value SPHT_K3 = 0.662

User