


LIMA\_03\_SV\_14\_SANDR

Company:  
 User:  
 Result file: C:\CamsizerXT\CAMDAT\LIMA\_03\_SV\_14\_SANDR\LIMA\_03\_SV\_14\_SANDR036.rdf  
 Task file: C:\CamsizerXT\CAMSYS\LIMA\_03\_SV\_14\_SANDR.afg  
 Time: 4/30/2015 , 12:47 , duration 5 min 57 s at 0.5 % covered area, image rate 1:1 , with X-Fall, gap width = 7.0 mm

Particle model: xc\_min  
 No. of particles: CCD-B = 803206 , CCD-Z = 16089  
 Velocity adaption: no  
 Fitting: no

Material:

ASTM(+)	No(-)	retained [%]	passing [%]	SPHT3	Symm3	b/l3	PDN
> #16		0.0	100.0				0
#20	#16	0.0	100.0				0
#30	#20	0.0	100.0				0
#40	#30	0.0	100.0				0
#50	#40	0.1	100.0	0.658	0.720	0.701	83
#60	#50	1.8	99.9	0.883	0.872	0.774	3880
#70	#60	18.3	98.1	0.921	0.902	0.775	57990
#140	#70	79.0	79.8	0.912	0.906	0.735	606053
PAN	#140	0.8	0.8	0.833	0.915	0.606	1849550



Mean value SPHT3 = 0.906  
 Mean value Symm3 = 0.903  
 Mean value b/l3 = 0.728  
 Mean value RDNS\_C3 = 0.614  
 Mean value SPHT\_K3 = 0.656

User