


SGVIR_01_SPHERE_AND_ROUND

Company:	
User:	
Result file:	C:\...zerXT\CAMDAT\SGVIR_01_SPHERE_AND_ROUND\SGVIR_01_SPHERE_AND_ROUND066.rdf
Task file:	C:\CamsizerXT\CAMSYS\SGVIR_01_SPHERE_AND_ROUND.afg
Time:	7/20/2015 , 15:55 , duration 9 min 8 s at 0.5 % covered area, image rate 1:1 , with X-Fall, gap width = 7.0 mm
Particle model:	xc_min
No. of particles:	CCD-B = 1582889 , CCD-Z = 25324
Velocity adaption:	no
Fitting:	no
Material:	

ASTM(+)	No(-)	retained [%]	passing [%]	SPHT3	Symm3	b/l3	PDN
> #16		0.0	100.0				0
#20	#16	0.0	100.0				0
#30	#20	0.0	100.0	0.812	0.827	0.815	1
#40	#30	0.0	100.0	0.658	0.790	0.794	2
#50	#40	0.1	100.0	0.675	0.760	0.709	51
#60	#50	0.7	99.9	0.845	0.855	0.760	1741
#70	#60	5.9	99.2	0.885	0.880	0.763	20865
#140	#70	90.2	93.3	0.891	0.900	0.726	1056093
PAN	#140	3.1	3.1	0.824	0.913	0.592	2954365



Mean value SPHT3 = 0.881
Mean value Symm3 = 0.897
Mean value b/l3 = 0.709
Mean value RDNS_C3 = 0.523
Mean value SPHT_K3 = 0.614

User