


SGVIR_04_SPHERE_AND_ROUND

Company:
 User:
 Result file: C:\...zerXT\CAMDAT\SGVIR_04_SPHERE_AND_ROUND\SGVIR_04_SPHERE_AND_ROUND073.rdf
 Task file: C:\CamsizerXT\CAMSYS\SGVIR_04_SPHERE_AND_ROUND.afg
 Time: 7/21/2015, 10:18, duration 9 min 38 s at 0.5 % covered area, image rate 1:1, with X-Fall, gap width = 7.0 mm

Particle model: xc_min
 No. of particles: CCD-B = 3547087, CCD-Z = 93704
 Velocity adaption: no
 Fitting: no

Material:

ASTM(+)	No(-)	retained [%]	passing [%]	SPHT3	Symm3	b/l3	PDN
> #16		0.0	100.0				0
#20	#16	0.0	100.0	0.840	0.861	0.697	2
#30	#20	0.1	100.0	0.711	0.866	0.575	7
#40	#30	0.0	99.9	0.838	0.874	0.717	6
#50	#40	0.1	99.9	0.730	0.792	0.716	55
#60	#50	0.8	99.8	0.845	0.856	0.760	2348
#70	#60	6.3	99.0	0.888	0.883	0.760	26362
#140	#70	87.8	92.7	0.894	0.899	0.728	1319802
PAN	#140	4.9	4.9	0.845	0.915	0.621	17586764



Mean value SPHT3 = 0.884
 Mean value Symm3 = 0.896
 Mean value b/l3 = 0.710
 Mean value RDNS_C3 = 0.527
 Mean value SPHT_K3 = 0.617

User