


VAL_01_SV_14_SPHERE_ROUND

Company:	
User:	
Result file:	C:\...izerXT\CAMDAT\VAL_01_SV_14_SPHERE_ROUND\VAL_01_SV_14_SPHERE_ROUND096.rdf
Task file:	C:\CamsizerXT\CAMSYS\VAL_01_SV_14_SPHERE_ROUND.afg
Time:	6/8/2015 , 16:31 , duration 7 min 9 s at 0.5 % covered area, image rate 1:1 , with X-Fall, gap width = 7.0 mm
Particle model:	xc_min
No. of particles:	CCD-B = 1854786 , CCD-Z = 28966
Velocity adaption:	no
Fitting:	no
Material:	

ASTM(+)	No(-)	retained [%]	passing [%]	SPHT3	Symm3	b/l3	PDN
> #16		0.0	100.0				0
#20	#16	0.0	100.0				0
#30	#20	0.0	100.0	0.777	0.850	0.818	2
#40	#30	0.0	100.0				0
#50	#40	0.1	100.0	0.698	0.755	0.731	88
#60	#50	1.5	99.9	0.872	0.872	0.769	4914
#70	#60	11.2	98.4	0.902	0.892	0.767	53901
#140	#70	84.5	87.2	0.895	0.902	0.729	1214155
PAN	#140	2.7	2.7	0.849	0.921	0.633	4543688



Mean value SPHT3 = 0.888
Mean value Symm3 = 0.900
Mean value b/l3 = 0.716
Mean value RDNS_C3 = 0.558
Mean value SPHT_K3 = 0.630

User